

M-83

AHD/TVM/DVC/TVM/181/2024

District Veterinary Centre
Thiruvananthapuram
E mail:dvctvm.ahd@kerala.gov.in
Date:18/02/2025

WINDOW NOTICE

1	Tender No& Superscription	Tender No 04/2024-25 Tender for Purchase of V Top OT Table
2	Name of Department	Animal Husbandry
3	Official Address	District Veterinary Centre,PMG,Thiruvananthapuram-695033
4	Details of Requirements	Purchase of V Top OT Table
5	Price of Tender form	Rs 500/- +12%GST Rs 500/- +60=610/-
6	EMD	1% of the total cost
7	Closing Date & Time for sale of Tender form	10/03/2025 5.00pm
8	Estimated cost	2Lakhs(Two Lakhs)
9	Date upto which the rates are to be firm for acceptance	31/03/2025
10	The last date for receipt of completed tenders	10/03/2025 5.00pm
11	Opening date of the tenders	11/03/2025 11.00am

Contact Number:04712302643

Place: Thiruvananthapuram
Date :18/02/2025

K C Prasath

Digitally signed by K C Prasath
DN: cn=K C Prasath, o=Kerala Veterinary, ou=Kerala Veterinary, email=K.C.Prasath@kerala.gov.in
Date: 2025.02.18 07:25:40+05:30
Reason: I am the issuer of the document
Faint P22 - Reader version: 2023.1.0



**Government of Kerala
Animal Husbandry Department**

DISTRICT VETERINARY CENTRE

THIRUVANANTHAPURAM

No : AHD/TVM/DVC/TVM/181/2024

18 February 2025

TENDER NOTICE

TENDER No.4/2024-2025

DISTRICT VETERINARY CENTRE, THIRUVANANTHAPURAM

**TENDER FOR PURCHASE OF V TOP OT TABLE FOR DISTRICT VETERINARY
CENTRE, THIRUVANANTHAPURAM**

Competitive tenders are invited from manufacturers and authorized suppliers under "TWO COVER SYSTEM" for Supply and Installation of **V top OT table** on a Plug and Play basis for the use of District Veterinary Centre, Thiruvananthapuram for the year 2024-25

First Cover (Cover A- Technical) will have following enclosures

1. The item intended to be quoted compared with each of the specification and requirement attached with this tender document and without rates. The equipment should be in strict accordance with our **specification**. The make, brand, model, catalogue number etc. must be clearly shown. The relevant literature/leaflets/catalogues/brochures of the equipment should accompany the documents.

2. The list of institutions where the equipment has been supplied and installed should be furnished.
3. Specification Compliance of the quoted item compared with each of the minimum requirement tendered.
4. The tenderer has to give in writing that they shall comply with all the general and special conditions, for purchase of equipment that are described in the tender.
5. The item should be quoted by Original Equipment Manufacturers (OEM's). If not the authorization letter from the OEM should be produced.
6. Attested copy of GST certificate.
7. Annual Turnover Certificate from Chartered Accountant for last three years.
8. Standard Certificates for the equipment and all the other relevant National or International certificates for the equipment if any.
9. Original tender form duly signed by authorized signatory.
10. Preliminary agreement in Kerala Stamp paper worth Rs.200/-.
11. An affidavit in stamp paper worth Rs.200/- notarised by a Notary Public of either the place where the person making the affidavit is residing or Thiruvananthapuram signed by the responsible person of firm (Managing Director / Director) specifying his status, age, residential address as well as his responsibility in the conduct of the business for the firm in the State.
12. A covering letter with name, address, pin code, Telephone/fax no., Email address and other details of the firm with the list of enclosures to be specified and addressed to the Chief Veterinary Officer, District Veterinary Centre, Thiruvananthapuram.
13. The Bank details of the firm for enabling payment by online transaction should be provided.
14. A declaration of the authorized sales and service agents /manufacturer making the supply of the materials to the effect that the supply of the materials will be done within 30 days from the date of receipt of the supply order. If any grace period to effect the supply, installation, successful demonstration and training is required by the bidder, it should be mentioned well beforehand with the quote. The contact details and phone numbers of the service engineers of the firm should essentially be provided.

15. Declaration of Onsite warranty for Three years for the V top OT table, ensuring the availability of expert service and maintenance, uninterrupted availability of spare parts and repair for next ten years must be assured.

Second cover (cover -B Financial) will have the following enclosures

1. Rate should be quoted as basic price, rate of tax and total amount per unit in BOQ
2. AMC for the next five years after successful completion of warranty period should be provided as a separate annexure.

Cover A and B will be opened as per the schedule given below. Cover B of those tenderers who satisfy the above conditions in Cover A will only be opened. Intending tenderers may sent the exact amount towards the cost of tender form online. Late tenders will not be considered under any circumstances. The tenderers whose cover "B" is opened should produce on demand, the original documents specified in Cover "A" for verification.

SCHEDULE OF SUBMISSION AND OPENING OF TENDER No. 4/2024-2025

1	Due date and time of submission of tenders (Cover A and B)	10/03/2025 05.00 PM
2	Date and time of opening of Cover A	11/03/2025 11.00 AM
3	Date and time of publishing list of Tenders eligible for opening their Cover B	Will be informed later
4	Date and time of opening of Cover B	Will be informed later
5	Date upto which the rates are to be firm for acceptance	31/3/2025
6	Price of tender form (Original)	Rs.500/- + GST 12 % Rs.500/- + 60 = Rs 610/-
7	Designation and Address of the Officer to whom the tenders are to be submitted	Chief Veterinary Officer, District Veterinary Centre, PMG, Thiruvananthapuram- 695033

Thiruvananthapuram



K C Prasath
Digitally signed by K C Prasath
DN: cn=K C Prasath, o=Thiruvananthapuram, ou=District Veterinary Centre, email=k.prasath@thiruvananthapuram.gov.in, c=IN
Reason: I am the author of this document.
Date: 2025.03.10 09:11:45 +05'30'
Foxit PDF Reader Version: 2024.2.0

Chief Veterinary Officer,
District Veterinary Centre,
Thiruvananthapuram

1. SPECIFICATION FOR THE SUPPLY AND INSTALLATION OF V-TOP
ELECTRIC VETERINARY OPERATING TABLE WITH TEMPERATURE
CONTROL

1. The table top and all fittings should be constructed of SS 316/304.
2. All non-SS fittings should be suitably treated to prevent corrosion and water/fluid retention. The coating provided should be resistant to liquids, acids and other corrosive agents.
3. A two-panel table top system that adjusts from horizontal to near vertical to hold the pets in the desired position during surgery.
4. V-top Swivel Angle: 0°- 90°, with secure locking mechanism and independently adjustable panels.
5. Overall size of the top should be at least 32"W× 60"L. (81.4cm W x 152.4 cm L).
6. Each panel should contain a heated area of at least 8" x 48", with an overall heated area of at least 16" x 48".
7. Controlled heating in temperature ranges from 70°F to 97°F with even heat distribution, with digital display and control.
8. Heating system completes automatic self-diagnostic test every time the unit is activated.
9. Electric lifting actuator with minimum holding capacity of 120kg (265 lbs) on the surgery top.
10. Hand remote control should be provided.
11. Actuators and other electrical components should be compatible with standard electrical fittings and voltages.
12. Removable SS Waste channel with drain hole to allow for easy non-spill drain flow.
13. Close fitting translucent bucket for waste collection with handle for easy removal.

14. The IV pole mount should be welded on and the pole adjustable to at least 30 inches height.
15. Provision should be available for attaching an instrument tray.
16. Table height should be electronically adjustable from 28" to 40" (70cm to 120 cm).
17. Table should be capable of Trendelenburg Positioning from flat to nearly vertical.
18. 3" Rubber Caster wheels with locks and brakes installed for smoother movement and positioning and easier braking.
19. Control box and pedal controls should be waterproof or shielded against fluid ingress, and positioned for easy accessibility and operation.
20. Should have provision at all four ends for patient restraints, mounted on sliding rails.
21. Table top and fittings should be fully electrically insulated/isolated from the heating elements, lifting mechanism and other structural components.
22. All exposed metal edges should be rolled or have seal welds without any sharp edges, to prevent injury and fluid ingress and corrosion.

SPECIAL CONDITIONS FOR SUPPLY AND INSTALLATION OF EQUIPMENTS

ON A PLUG AND PLAY BASIS - 2024-25

1. The rates quoted should be inclusive of all taxes and should be quoted in Indian currency only.
2. Firms blacklisted in Kerala State or other States cannot participate in the tender.
3. The Security Deposit should be furnished as Account Payee Demand Draft/ Bank Guarantee from a Nationalised Bank or Treasury Deposit and will be released only after satisfactory completion of warranty period subsequent to production of all preventive maintenance documents.
4. The rates should be for delivery F.O.R. at District Veterinary Centre, Thiruvananthapuram and should include cost towards Loading and unloading, transit insurance, installation charges and any other charges as applicable.
5. Ex-godown rates and F.O.R station of Despatch rates offered by the firms are not acceptable and such tenders will be summarily rejected.
6. Successful tenderers should supply the items at the approved rates and conditions to District Veterinary Centre, Thiruvananthapuram.
7. The successful tenderers should complete all supplies within the stipulated time stated in the supply orders.
8. The non-availability of materials, price hike of materials, transportation problems etc will not be accepted as reasons for non-supply or late supply.
9. The only jurisdiction for any dispute regarding this purchase will be at Thiruvananthapuram.
10. If any of the date scheduled for opening of tender is declared as holiday (local or general), the opening will be on the next working day at the same time.
11. Name and place to which the items are to be supplied is "Office of the Chief Veterinary Officer, District Veterinary Centre, Thiruvananthapuram, Kerala".
12. The supplies should be effected only by the original tenderer and the invoice should also be furnished only by them.
13. The Chief Veterinary Officer, District Veterinary Centre, Thiruvananthapuram reserves the right to delete, waive any of the above special Conditions or cancel the whole tender at any stage of tender without assigning any reason thereof.

14. The supply should be completed within 30 days from the date of supply order.
15. Equipment should be tested and certified at the time of installation
16. The equipment should come with all standard fittings, accessories and spares, if any in a plug and play basis
17. Full payment will be made by the office, after the receipt of equipment unit(s) in good condition after installation, demonstration and satisfactory functioning. No advance payments will be made.
18. The equipment/machinery/items should be in strict accordance with our specification. If the offered equipment does not fulfill any of the specifications, this has to be clearly mentioned in the bid. If any of the specifications is omitted in the bid without mentioning it, the offer will be treated as confirming to our specification. The make, brand etc. must be clearly shown in the tender itself. The relevant literature/leaflets/catalogues/brochures of the equipment should accompany the tender documents
19. The guarantee period, availability of spare parts, after sales service facility etc. are to be clearly mentioned in a separate sheet. A certificate to the effect that spares of the equipment quoted will be made available at least for ten years from the date of installation should be submitted along with the tender. The equipment should have a minimum warranty period of Three years from the date of installation.
20. The rates for Annual Maintenance Contract for a period of five years after the warranty period should also be submitted along with the commercial bid. The full details (address and contact details) of the firm/ firms and personnel undertaking coverage of warranty period and services thereafter should be mentioned as a separate annexure
21. The manufacturer should impart operational training to technicians of the hospital. The successful supplier should demonstrate the performance of the equipment at site.

Thiruvananthapuram

K C
Prasath

Digitally signed by K C Prasath
DN: cn=K Prasath, o=K Prasath, ou=K Prasath, email=k.prasath@kprasath.com, c=IN
Reason: I am the author of the document
Date: 2025.02.10 10:12:25 +05'30'
File: PDF Signature Version: 2.0.0

Chief Veterinary Officer,
District Veterinary Centre,

